Indirect Calorimetry ICSLA_CAL_003

Purpose

Indirect calorimetry provides detailed information on the energy metabolism of mutant mice. Energy expenditure is evaluated through indirect calorimetry by measuring oxygen consumption with an open flow respirometric system. CO_2 and O_2 sensors measure the difference in CO_2 and O_2 concentrations in air volumes flowing through control or animal cages. The amount of oxygen consumed over a given period of time can thus be calculated, as far as the air flow through the cage is known. Data are expressed as ml O_2 h⁻¹animal⁻¹. The system also monitors CO_2 production, therefore, the respiratory exchange ratio (RER) and heat production can be calculated. An activity and food and water intake monitoring system can also be integrated into the set up in order to investigate circadian pattern and behaviour.

Ontological description: MP:0005266 - abnormal metabolism.

Experimental Design

Minimum number of animals: 7M or 7F

Age at test: Week 53

Sexual dimorphism: In general, female mice have higher metabolism compared to males therefore statement is not entirely correct. However, genotype x sex interaction are rare therefore testing only males is acceptable.

It is essential that all phenotyping experimentation is conducted at the same time of day because physiological and biochemical parameters e.g. metabolic rate, body temperature and activity are subject to temporal rhythms. In the indirect calorimetry module standard measurements begin five hours before lights-off (lights off = T0) and are finished at T16 i.e. four hours after lights-on the next morning. Optional: Mice can be given one day of acclimation before the trial, and the trial can be continued for more than 21 hours.

Equipment

- 1. Calorimetric system equipped with respirometer, feeder and water bottles
- 2. Ambulatory activity monitor (dependent on system specifications)
- 3. Food and water intake monitor
- 4. Computer with apparatus software installed

Procedure

- 1. Optionally mice are allowed to acclimatise to the phenotyping room, to the calorimetry cage, food hoppers and drinking bottles 24 hours before testing.
- **2.** Prepare and calibrate the calorimetric apparatus to confirm the accuracy of the gas sensors and flow meters. Specifically prior to each experiment:
- **a.** Apply known volumes of CO₂ and O₂ to determine the sensitivity of the gas sensors and flow meters.
- b. Run a complete calibration protocol according to the manufacturer's recommendations.
- **3.** Provide each calorimetry cage with sufficient food and water for a period of ~24 (or 48) hours.
- **4.** Weigh the mouse.
- 5. Place the mouse into a calorimetry cage with food and water available ad libitum.
- **6.** Label the chamber with the corresponding subject identification and close it ensuring there is adequate air flow.
- 7. Initiate the calorimetric system for measurement:
- **a.** Set up a new experiment in accordance with the manual (or load a file from a previous experimental setting).
- **b.** Start recording measurements five hours before lights off for a total duration of 21 hours at minimum. Optional: 24 hours acclimation can be applied and the recording may continue for 48 hours.
- **c.** The latency of CO₂ and O₂ activity transmitted and recorded is dependent on the number of chambers in use but will be logged periodically.
- **8.** Generating a data report:
- a. Upload all data from the experimentation including:
- Gas analysis VO2 and VCO2 (ml/h/animal)
- Heat production (kJ/h/animal)
- Periodicity of measurements taken throughout experimentation (Figure 1)
- Animal and the corresponding chamber that was used
- The respiratory exchange ratio (RER) can be calculated using the VCO2/VO2 ratio.
- **9.** Activity parameters recorded will depend on the specification of calorimetric system used:
- a. Ambulatory activity can be derived from the number beam splits during the session
- **b.** Total activity can be derived from the number of fine movement (e.g. grooming behaviours) as well as ambulatory activity
- **c.** An average of each of these parameters of activity is calculated hourly across the measurement period (between T-5 and T16).
- **d.** Water and food intake (cumulative, hourly or total food and water intake, between T-5 and T16, will be computable depending on the calorimetric system used).

- 10. Remove each mouse from its chamber in turn at the end of the experimental session and record its weight. Return to their home cage.
- 11. Monitor the animals carefully to observe any abnormal behaviour(s). Ensure that food and water are available *ad libitum*.
- 12. Wash and wipe clean the chambers with warm water and dilute alcohol or appropriate disinfectant respectively.




Figure 1. Daily workflow of calorimetric experimentation (Note: T0 designates start of dark cycle).

Notes

The system requires periodic calibration of the gas sensors and flow meters to ensure precise measurements. The calibration procedure consists of the application of a gas of known composition and adjusting control knobs in the front of the oxygen and carbon dioxide sensors to obtain readings that reflect the contents of the calibration gas. It is recommended that the system be calibrated prior to the start of each experiment. The analyzers should not be shut down if not urgently required for maintenance. If this has to be done a warm up time of at least 90 minutes is required for the gas sensors for calibration (refer to manufacturer's manual). Calibrations and shut downs should be recorded in the laboratory journal.

Calorimetry test is to be performed before the ECG/ECHO test to avoid effects of hair removal on the calorimetry results.

The information about the date of the experiment, that is the date when the measurement is performed, is an important parameter which is to be submitted in the Experiment xml file (dateOfExperiment="2013-02-28").

Data QC

- 1. Respiratory Exchange Rate (RER) is between 0.7-1.00
- 2. Mice show normal feeding and drinking behaviour
- 3. Mice show stable weight before and after calorimetry
- 4. Correct calibration of gases according to manufacturer's manual

MetaData and examples

Metadata	Example
Time of dark cycle start	The starting time of the dark cycle. E.g. 7 pm
Time of dark cycle end	The ending time of the dark cycle. E.g. 7 am
Room temperature	The range of min. and max. temperatures of the calorimetry chamber. If the calorimetry machine doesn't register the chamber temperature, the room temperature in which the machine resides can be reported instead. E.g. 20.0-24.0 (C°). If the temperature is constant throughout the experiment, a single temperature can be submitted. E.g. 20.0 (C°). Do not submit the subtraction between the 2 values!
Acclimation to calorimetry cages	E.g. Yes/No.
Duration of test	Duration of the test without including the acclimation period. Can be a minimum of 21 hours or more, when acclimation is done. E.g. 21 (hours).
Equipment ID	ID of the machine used when more than 1 is used having same model and manufacturer. E.g. machine 1, machine 2, machine Minnie, machine Mickey Mouse, etc.
Equipment manufacturer	Manufacturer of the equipment. E.g. TSE Systems GmbH.
Equipment model	Model of the equipment. E.g. Labmaster CaloSys.
Experimenter ID	An ID of any format to be used coherently both inside the same procedure and for all procedures. E.g. Harw_001, or 1/2/3.
Date equipment last calibrated	Most recent date in which the equipment (or any part of) used in the procedure was subject to a calibration event.
Outer dimension of cage	(L, cm) x (W, cm) x (H, cm). E.g. 21.4 x 11.5 x 13.3

Inner dimension of cage	(L, cm) x (W, cm) x (H, cm). E. g. 21.4 x 11.5 x 13.3
Height from platform** to lid assembly	cm. E.g. 8.6
Available space for mouse	(L, cm) x (W, cm) x (H, cm). E.g. 20.3 x 10.4 x 8.6
Infrared beam setting on X axis	Orizontal activity: locomotion. N beams with L cm spacing, H cm above platform**. E.g. 16, 1.3, 3.0
Infrared beam setting on Y axis	Orizontal activity: locomotion. N beams with L cm spacing, H cm above platform**. E.g. 16, 1.3, 3.0
Infrared beam setting on Z* axis	Vertical activity: rearing/jumping. N beams with L cm spacing, H cm above platform**. E.g. 16, 1.3, 6.8
Beam strip placement on exterior of chamber****	Left/right, front/back, both
Lightbeam wavelength	nm, E.g. 950
Scanning rate	Hz, E.g. 100
Refractory period	s, E.g. 0.8
Presence of bedding into the cage	Yes/No
Igloo in cage	Yes/No

^{*}Z has same direction as X only is higher with respect to the platform

Parameters and Metadata

Room temperature max ICSLA_CAL_023_001 | v1.4

^{**} the floor where the mouse stand

^{***} This value is the "Height from platform** to lid assembly"

procedureMetadata

Req. Analysis: false Req. Upload: true Is Annotated: false

Unit Measured: C

Acclimation to respirometry cages ICSLA_CAL_012_001 | v1.0

procedureMetadata

Req. Analysis: false Req. Upload: true Is Annotated: false

Options: Yes, No,

.....

Equipment ID ICSLA_CAL_014_001 | v1.0

procedureMetadata

Req. Analysis: false Req. Upload: true Is Annotated: false

Duration of test ICSLA_CAL_013_001 | v1.0

procedureMetadata

Req. Analysis: false Req. Upload: true Is Annotated: false

Unit Measured: Hours

Calibration method procedureMetadata	CSLA_CAL_040_001 v1.	0	
Req. Analysis: true	Req. Upload: false	Is Annotated: false	
Options: Gas bottle method,	Reference range method,		
Refractory period ICSLA_CAL_037_001 v1.0 procedureMetadata			
Req. Analysis: true	Req. Upload: false	Is Annotated: false	
Unit Measured: S			
Time of dark cycle start ICSLA_CAL_010_003 v3.0 procedureMetadata			
Req. Analysis: false	Req. Upload: true	Is Annotated: false	

Options: 21, 48, 22, 23,

Beam strip placement on exterior of chamber ICSLA_CAL_034_001 | v1.0

procedureMetadata

Reg. Analysis: true Reg. Upload: false Is Annotated: false Options: Front/back, Left/right, Both, Inner dimension of cage ICSLA_CAL_028_001 | v1.0 procedureMetadata Req. Analysis: true Req. Upload: false Is Annotated: false Unit Measured: cm Body weight after experiment ICSLA_CAL_002_001 | v1.3 simpleParameter Req. Analysis: false Req. Upload: true Is Annotated: false Unit Measured: q

Respiratory Exchange Ratio series ICSLA_CAL_042_001 | v1.0

Req. Analysis: false	Req. Upload: false	Is Annotated: false
Increments: Minimum 21		
Equipment manufa procedureMetadata	cturer ICSLA_CAL_015_0	001 v1.1
Req. Analysis: true	Req. Upload: true	Is Annotated: false
Options: O'hara Co. Ltd., Sab	le Systems, Columbus Instrum	ents, TSE Systems GmbH,
Body weight before simpleParameter	e experiment ICSLA_C	GAL_001_001 v1.3
Req. Analysis: false	Req. Upload: true	Is Annotated: false
Unit Measured: g		
Total activity (no. o LA_CAL_007_001 v1.2 seriesParameter	of fine movement +	no. of beam cuts) ics

Req. Analysis: false Req. Upload: false Is Annotated: true

Unit Measured: count/hour		
Increments: Minimum 21		
Room temperature procedureMetadata	min ICSLA_CAL_024_001	I v1.2
Req. Analysis: false	Req. Upload: true	Is Annotated: false
Unit Measured: C		
Infrared beam setti	ng on Y axis ICSLA_C	ML 022 004 Lv4 0
procedureMetadata	ing off i axis lostA_c	AL_032_001 V1.0
Req. Analysis: true	Req. Upload: false	Is Annotated: false
Unit Measured: cm		
Presence of beddir	ng into the cage ICSL	_A_CAL_038_001 v1.0
Pog Analysis: truo	Pog Upload: folco	le Annotated: falco
neq. Analysis. Hue	Req. Upload: false	Is Annotated: false
Options: Yes, No,		

Carbon dioxide production ICSLA_CAL_004_001 | v1.2

seriesParameter

Req. Analysis: false Req. Upload: true Is Annotated: true

Unit Measured: ml/h/animal

Increments: Minimum 21

Cumulative water intake ICSLA_CAL_022_001 | v1.1

seriesParameter

Req. Analysis: false Req. Upload: false Is Annotated: true

Unit Measured: ml

Increments: Minimum 21

.....

Oxygen consumption ICSLA_CAL_003_001 | v1.2

seriesParameter

Req. Analysis: false Req. Upload: true Is Annotated: true

Unit Measured: ml/h/animal

Increments: Minimum 21

.....

Acclimation period ICSLA_CAL_041_001 | v1.0

procedureMetadata

Req. Analysis: false Req. Upload: false Is Annotated: false

Unit Measured: Hours

.....

Respiratory Exchange Ratio ICSLA_CAL_017_001 | v1.2

simpleParameter

Req. Analysis: false Req. Upload: false Is Annotated: true

Derivation:

div(meanOfIncrements('ICSLA_CAL_004_001',21), meanOfIncrements ('ICSLA_CAL_003_001',21))

.....

Equipment model ICSLA_CAL_016_001 | v1.2

procedureMetadata

Req. Analysis: true Req. Upload: true Is Annotated: false

Options: Oxymax FAST, PhenoMaster/Labmaster CaloSys, (Dri)+Fed, Act.X,Y 12Mice,

Oxymax/CLAMS, FWI-3002 & IA-16M,

PhenoMaster/Labmaster CaloSys, (Dri)+Fed, Act.X,Y 20Mice,

PhenoMaster/LabMaster CaloSys, (Dri)+Fed, Act.X,Y 16mice, SAMPLEMAX,

SM-MARS 8 channel Metabolic system,

PhenoMaster/Labmaster CaloSys, (Dri)+Fed, Act.X,Y 32Mice, PhenoMaster/Labmaster CaloSys, (Dri)+Fed, Act.X,Y 12Mice - Gas system 994620 series,		
Total food intake IC simpleParameter	SLA_CAL_008_001 v1.2	
Req. Analysis: false	Req. Upload: false	Is Annotated: true
Unit Measured: g		
Available space for procedureMetadata	r mouse ICSLA_CAL_03	0_001 v1.0
Req. Analysis: true	Req. Upload: false	Is Annotated: false
Unit Measured: cm		
Outer dimension of procedureMetadata	f cage ICSLA_CAL_027_0	001 v1.0
Req. Analysis: true	Req. Upload: false	Is Annotated: false
Unit Measured: cm		

Infrared beam setting on Z axis ICSLA_CAL_033_001 | v1.0

procedureMetadata

Req. Analysis: true Req. Upload: false Is Annotated: false Unit Measured: cm Infrared beam setting on X axis ICSLA_CAL_031_001 | v1.0 procedureMetadata Req. Analysis: true Req. Upload: false Is Annotated: false Unit Measured: cm Chamber temperature min ICSLA_CAL_025_001 | v1.1 procedureMetadata Req. Analysis: false Req. Upload: false Is Annotated: false Unit Measured: C

Req. Analysis: true **Req. Upload:** false **Is Annotated:** false Unit Measured: nm Igloo in cage ICSLA_CAL_039_001 | v1.0 procedureMetadata Reg. Analysis: true Reg. Upload: false Is Annotated: false Options: Yes, No. Chamber temperature max ICSLA_CAL_026_001 | v1.1 procedureMetadata Reg. Analysis: false Reg. Upload: false Is Annotated: false Unit Measured: C

Heat production (metabolic rate) ICSLA_CAL_005_001 | v1.2

seriesParameter

Req. Analysis: false Req. Upload: true Is Annotated: true

Unit Measured: kJ/h/animal

Increments: Minimum 21		
Experimenter ID ICS procedureMetadata	SLA_CAL_018_001 v1.0	
Req. Analysis: false	Req. Upload: true	Is Annotated: false
Cumulative food in	ntake ICSLA_CAL_009_00	1 v1.2
seriesParameter		
Req. Analysis: false	Req. Upload: false	Is Annotated: true
Unit Measured: g		
Increments: Minimum 21		
Ambulatory activity (no. of beam cuts) ICSLA_CAL_006_001 v1.2 seriesParameter		
Req. Analysis: false	Req. Upload: false	Is Annotated: true
Unit Measured: count/hour		
Increments: Minimum 21		

Time of dark cycle end ICSLA_CAL_020_003 v3.1 procedureMetadata			
Req. Analysis: false	Req. Upload: true	Is Annotated: false	
Scanning rate ICSLA	CAL 036 001 Lv1 1		
procedureMetadata	_0/12_000_001 V1.1		
Req. Analysis: true	Req. Upload: false	Is Annotated: false	
Unit Measured: Hz			
Height from platfor procedureMetadata	m to lid assembly	CSLA_CAL_029_001 v1.0	
Req. Analysis: true	Req. Upload: false	Is Annotated: false	
Unit Measured: cm			

Date equipment last calibrated ICSLA_CAL_019_001 | v1.2

procedureMetadata

Req. Analysis: false	Req. Upload: false	Is Annotated: false
Total water intake	ICSLA_CAL_021_001 v1.1	
Req. Analysis: false	Req. Upload: false	Is Annotated: true
Unit Measured: ml		
