

Electrocardiogram (ECG) ICSLA_ECG_002

Purpose

To provide a high throughput method to obtain Electrocardiograms in a conscious mouse or an anesthetized mouse.

Experimental Design

- **Minimum number of animals** : 5M + 5F
- **Age at test**: Week 54
- **Sex**: We would expect the results of this test to show sexual dimorphism

Procedure

Conscious Procedure with ECGenie

1. The lead plates are to be snapped into place onto the top of the pre-amplifier tower. The covering is removed to reveal three gel coated pads surrounded by a sticking plate. The plate will need to be covered with the extra cover in the package.
2. Turn on the combined amplifier and the pre-amplifier tower.
3. Double click the icon ECG acquisition on the acquisition computer.
4. Open the ECG set up file (for default settings).
5. Place mouse on pad, lowering the Red Acrylic Cubby to surround the mouse on 3 sides discouraging escape.
6. Press Start.
7. After the desired acquisition time, (5-10 minutes) stop the reading. There will be one long reading.
8. Save the data.
9. For additional readings create a new session using the same settings as before.
10. When saving sections with good readings, highlight the selected area and then save.

Anesthetized Procedure

1. Place mouse in the induction chamber. Anesthetize with 2.5-4% isoflurane in oxygen.
2. Transfer mouse to a warmed platform and maintain with a nose cone at 2-2.5% isoflurane.
3. Fasten the mouse to the heated platform set to 37.5-39° Add electrode cream if using the foot surface electrodes on the platform surface. *Alternatively insert needle electrodes subcutaneously into the limbs using Lead Configuration I: Left Arm – Right Arm (LA-RA) and/or Lead Configuration II: Left Leg – Right Arm (LL-RA).*
4. Monitor the body temperature with a rectal probe if possible.

5. Open the recording software with appropriate settings and record an ECG for about 2 minutes.
6. Save the recording and allow mouse to recover.

Notes

Data Analysis - Conscious Procedure with ECGenie

1. Open Emouse Analyses icon
2. Select ECG signals
3. Choose folder (all readings in folder will show)
4. Click PNN X (for mice: N-N> than 6 ms)
5. Choose file(s) by highlighting
6. Go
7. Bottom file is the corrected file
8. Red dots should be on peak of R waves, if image appears inverted click invert
9. Click Add, or minus if R waves are not marked with red dots or if too many are marked
 - L click to zoom in
 - R click to zoom out
11. "What if?" button to remove unwanted sections
 - L click image (zooms in)
 - L click left boundary
 - L click right boundary
13. Options- click more if want to exclude more sections
14. Undo available
15. Go
16. Here can input animal data if desired
17. Save- For the first mouse in in group, hit save, a new results folder will be created within the folder with the mouse data. Then can click quick save or next.
18. For the rest of the mice in the series, can hit quick save at this point- saves in last selected file – will group all files together in same excel sheet.
19. Open Emouse Analyses icon
20. Select ECG signals
21. Choose folder (all readings in folder will show)
22. Click PNN X (for mice: N-N> than 6 ms)
23. Choose file(s) by highlighting
24. Go
25. Bottom file is the corrected file
26. Red dots should be on peak of R waves, if image appears inverted click invert
27. Click Add, or minus if R waves are not marked with red dots or if too many are marked
 - L click to zoom in
 - R click to zoom out
29. "What if?" button to remove unwanted sections
 - L click image (zooms in)
 - L click left boundary
 - L click right boundary
31. Options- click more if want to exclude more sections
32. Undo available
33. Go
34. Here can input animal data if desired

35. Save- For the first mouse in in group, hit save, a new results folder will be created within the folder with the mouse data. Then can click quick save or next
36. For the rest of the mice in the series, can hit quick save at this point- saves in last selected file - will group all files together in same excel sheet

Examples of good readings



Data QC

Analysis room should be dim and quiet. Keep the door closed preferably while analysis is taking place.

Figure A. Taking a reading




Figure B. Saving a section of the reading




Figure C. Analysis phase, with the options to remove sections on the “What if?” button below.



Data Analysis - Anesthetized Procedure


1. Review the tracing and note any abnormal findings. Take images for upload where required.
2. Analyse all or a selection of the tracing (approximately 100 beats minimum).
3. If using LabChart software (ADInstruments), select Mouse preset and QTc = Bazett.
4. The software can automatically mark the P, QRS and T waves of each beat (Beats to average = 1) and calculate the interval and amplitude data. The average of all beats will be determined for each parameter.
5. Alternatively, an averaged tracing can be generated first (Block Averaging). Markers for the P, QRS and T waves are placed by the software but can be adjusted according to the centre criteria. Interval and amplitude data is then generated by the software.

Example of the LabChart8 ECG setting



Example of LabChart ECG Averaging View

LabChart ECG Averaging View



Parameters and Metadata

CV ICSLA_ECG_003_001 | v1.0

simpleParameter

Req. Analysis: false

Req. Upload: false

Is Annotated: true

Unit Measured: %

HR ICSLA_ECG_002_001 | v1.1

simpleParameter

Req. Analysis: false

Req. Upload: true

Is Annotated: true

Unit Measured: bpm

Equipment ID ICSLA_ECG_016_001 | v1.0

procedureMetadata

Req. Analysis: false

Req. Upload: true

Is Annotated: false

Mean R amplitude ICSLA_ECG_013_001 | v1.1

simpleParameter

Req. Analysis: false

Req. Upload: false

Is Annotated: false

Unit Measured: mV

Anesthetic ICSLA_ECG_019_001 | v1.0

procedureMetadata

Req. Analysis: true

Req. Upload: true

Is Annotated: false

Options: Isoflurane, No anesthesia, Tribromoethanol, Avertin,

QTc Dispersion ICSLA_ECG_011_001 | v1.0

simpleParameter

Req. Analysis: false Req. Upload: false Is Annotated: false

Unit Measured: ms

Number of signals ICSLA_ECG_001_001 | v1.2

simpleParameter

Req. Analysis: false Req. Upload: true Is Annotated: false

ST ICSLA_ECG_008_001 | v1.0

simpleParameter

Req. Analysis: false Req. Upload: false Is Annotated: true

Unit Measured: ms

PQ ICSLA_ECG_005_001 | v1.0

simpleParameter

Req. Analysis: false Req. Upload: false Is Annotated: true

Unit Measured: ms

QRS ICSLA_ECG_007_001 | v1.2

simpleParameter

Req. Analysis: false

Req. Upload: true

Is Annotated: true

Unit Measured: ms

PR ICSLA_ECG_006_001 | v1.1

simpleParameter

Req. Analysis: false

Req. Upload: false

Is Annotated: true

Unit Measured: ms

HRV ICSLA_ECG_010_001 | v1.0

simpleParameter

Req. Analysis: false

Req. Upload: false

Is Annotated: true

Unit Measured: bpm

Noise level ICSLA_ECG_021_001 | v1.0

procedureMetadata

Req. Analysis: false Req. Upload: false Is Annotated: false

Waveform Image Comment ICSLA_ECG_026_001 | v1.0

simpleParameter

Req. Analysis: false Req. Upload: false Is Annotated: false

rMSSD ICSLA_ECG_014_001 | v1.0

simpleParameter

Req. Analysis: false Req. Upload: false Is Annotated: true

Unit Measured: ms

pNN5(6>ms) ICSLA_ECG_015_001 | v1.2

simpleParameter

Req. Analysis: false Req. Upload: false Is Annotated: false

Unit Measured: %

RR ICSLA_ECG_004_001 | v1.2

simpleParameter

Req. Analysis: false

Req. Upload: true

Is Annotated: true

Unit Measured: ms

Equipment Model ICSLA_ECG_018_001 | v1.0

procedureMetadata

Req. Analysis: true

Req. Upload: true

Is Annotated: false

Options: ECGenie, ECGenie + gel pads, Iso-DAM8A, ML866, ML870/p, ML826/FE132, PowerLab: 4/35,

Waveform Image ICSLA_ECG_025_001 | v1.0

seriesMediaParameter

Req. Analysis: false

Req. Upload: false

Is Annotated: false

Increments: Minimum 1

Mean SR amplitude ICSLA_ECG_012_001 | v1.1

simpleParameter

Req. Analysis: false Req. Upload: false Is Annotated: false

Unit Measured: mV

Light level ICSLA_ECG_022_001 | v1.0

procedureMetadata

Req. Analysis: false Req. Upload: false Is Annotated: false

Date equipment last calibrated ICSLA_ECG_023_001 | v1.1

procedureMetadata

Req. Analysis: false Req. Upload: false Is Annotated: false

Experimenter ID ICSLA_ECG_020_001 | v1.0

procedureMetadata

Req. Analysis: false Req. Upload: true Is Annotated: false

Equipment Manufacturer ICSLA_ECG_017_001 | v1.0

procedureMetadata

Req. Analysis: true Req. Upload: true Is Annotated: false

Options: AD Instruments, Mouse Specifics, Inc., World Precision Instruments,

QTc ICSLA_ECG_009_002 | v2.0

simpleParameter

Req. Analysis: false Req. Upload: true Is Annotated: false

Unit Measured: ms

Analysis Software ICSLA_ECG_024_001 | v1.0

procedureMetadata

Req. Analysis: true Req. Upload: false Is Annotated: false

Options: eMouse, Matlab, LabChart,

Lead Configuration ICSLA_ECG_027_001 | v1.0

procedureMetadata

Req. Analysis: false Req. Upload: false Is Annotated: false

Options: Lead I: LA-RA, Lead II: LL-RA,
